

Serial No. C 48689

**Weightron Bilanciai Ltd,**  
Titan Works  
Bridge Way off Broombank Road,  
Chesterfield Trading Estate,  
Chesterfield S41 9QJ, UK  
Tel: 01246 260062  
Fax: 01246 260844  
www.weightron.com



**Weightron Bilanciai Ltd,**  
Unit 4b, Block 4  
Westpark, Chelston  
Wellington TA21 9AD, UK  
Tel: 01823 662355  
Fax: 01823 662017  
www.weightron.com



## CALIBRATION CERTIFICATE

*This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL*

**Customer**

PH HULL + Sons Ltd  
Wolds farm  
SIX KILLS (AVE)  
ROGDALE  
LOCATION 6614 3PP.  
Site

**Machine Details**

Cust ref	22218
Make	Grant SDx
Model	GSE 562
Ser No	410339
Capacity	50,000 kg
e/d	13 kg
Working cap	

**Test Weights**

Set No.	Cert No.	Date Cal
1-27	6655	23/11/17
Cert Interval 12 mths		
Next Cert	Jan	2019

Tolerance +/-	
0-10000	2
10000-40000	40
40000-50000	60

Comments *all wts in kg.*  
 $R_u = 4640$   
 $T_u = 19360$   
*spread of 20000 kg.*

2
9  
3

**AS FOUND LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
10000	10000	0	400	0
<del>20000</del>	19960	-40	9980	-20
24640	24600	-40	19960	-40

**DEFINITIVE LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
10000	10000	0	400	0
20000	20000	0	10000	0
24640	24640	0	20000	0
19360	19360	0	19360	0
29360	29360	0	29360	0
39360	39360	0	39360	0
44000	44000	0		

**Repeatability**

Load	Ind	Ind	Ind

**Repeatability**

Load	Ind	Ind	Ind
39360	39360	39360	39360

**Eccentric Load**

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

**Eccentric Load**

Reading	Dev	Reading	Dev
1	10000	2	10000
3	10000	4	10000
5		6	
7		8	

Signature *[Signature]* Name *D Perkins* Date *09/01/2018*