

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of
Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PH HULL + SONS LTD
WOLDS FARM
SIX HILLS ROAD
RAGDALE
MILTON MANSFORD
LOCATION LE 14 3PP
Site

Machine Details

Cust ref	92378
Make	GSE
Model	562
Ser No	410339
Capacity	50000 kg
e/d	12 kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
pd 2	7251	02/12/18

Cert Interval	12 mths
Next Cert	JAN 2020

Tolerance +/-	Kg
0-10000	2
10000-40000	4
40000-50000	6

Comments
 S.H. = 4660
 T.H. = 19520

2	4
1	3

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
10000	10000	0	400	0
20000	20000	0	10000	0
24680	24680	0	20000	0
19520	19520	0	19520	0
29520	29520	0	29520	0
39520	39500	+20	39500	-20
44200	44180	-20		

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44200	44200	44180	44200

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 6680

Reading	Dev	Reading	Dev
1 6680	0	2 6680	0
3 6660	-20	4 6680	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature D. Davies Name D. Davies Date 05th JAN 2019