

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PK HULL & SONS LTD
WOLDS FARM
SIX MILLS ROAD
RACDALE
MELTON MOWBRAY
LEICESTERSHIRE
LOCATION LE14 3PA
IN/OUT

Machine Details

Cust ref	101755
Make	TNENT
Model	CSE 560
Ser No	410339
Capacity	50000kg
e/d	20kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1-22	7854	00119

Cert Interval	ANNUUM
Next Cert	JAN 2021

Tolerance +/-	kg
0-10000	20
10000-20000	40
20000-30000	60

Comments All weights in use
Fork lift - 4660
TEST UNIT - 19540

2	3
1	4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25540	25540	0	25540	0
29540	29540	0	29540	0
35540	35540	0	35540	0
39540	39540	0	39540	0
44200	44200	0	44200	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44200	44200	44200	44200

Repeatability

Load	Ind	Ind	Ind

Eccentric Load - 6660

Reading	Dev	Reading	Dev
1 6660	0	2 6660	0
3 6660	0	4 6660	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name Chris Wilson Date 10-1-20